VERIFICATION OF CALIBRATION REPORT
of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100672   Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program   Date:04/27/2012

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.73: 0.079
LOT #: 02511080A3
EXPIRATION: 02/01/2013

BLANK TEST	0.000	19:52
INTERNAL STANDARD	VERIFIED	19:52
EXTERNAL STANDARD	0.079	19:53
BLANK TEST	0.000	19:54
EXTERNAL STANDARD	0.078	19:54
BLANK TEST	0.000	19:55
EXTERNAL STANDARD	0.079	19:55
BLANK TEST	0.000	19:56
EXTERNAL STANDARD	0.079	19:56
BLANK TEST	0.000	19:57
EXTERNAL STANDARD	0.079	19:57
BLANK TEST	0.000	19:58

Average = 0.0788
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 2.00
PIC: 3.03
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C		PASSED
Breath Tube = 48.1°C		PASSED

PUMP INFO
Flow Rate = 5.028 L/M   PASSED

DETECTOR INFO
PUMP ON   PASSED
PUMP OFF   PASSED

FILTER INFO
Filter 1   PASSED
Filter 2   PASSED
Filter 3   PASSED

INTERNAL STANDARD   PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz   5/31/2012
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of May, 20 12

Nikki Roth, Notary Public
My Commission Expires With Office